



Quality Manual

Quality Manual

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 1 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		

AMERICAN DIAGNOSTIC CORP. • 55 Commerce Drive, Hauppauge, NY 11788 • Voice: 631-273-9600 • Fax: 631-273-9659 • email: info@adctoday.com



Quality Manual

Section	Point	TITLE	Page
1		Quality Management System	N/A
1	1.1	General	4
1	1.1.5	Process Flowchart	5
1	1.2	Documentation	6
1	1.3	Quality Manual	7
1	1.4	Control of Documents	8
1	1.5	Control of Quality Records	9
2		Management Responsibility	N/A
2	2.1	Management Commitment	9
2	2.2	Customer Focus	10
2	2.3	Quality Policy	10
2	2.4	Planning	11
2	2.5	Responsibility, Authority, and Communication	12
2	2.6	Management Review	14
3		Resource Management	N/A
3	3.1	Provision of Resources	15
3	3.2	Human Resources	15
3	3.3	Infrastructure	16
3	3.4	Work Environment	16
4		Product Realization	N/A
4	4.1	Planning of Product Realization	17
4	4.2	Customer Related Processes	18
4	4.3	Design and Development	20
4	4.4	Purchasing	22
4	4.5	Product and Service Provision	23
4	4.6	Cleanliness of Product and Contamination Control	24
4	4.7	Installation Activities	24
4	4.8	Servicing Activities	25
4	4.9	Particular Requirements for Sterile Medical Devices	25
4	4.10	Validation of Processes for Product and Service Provision	25
4	4.11	Identification and Traceability	26
4	4.12	Customer Property	27
4	4.13	Preservation of Product	27
4	4.14	Control of Monitoring and Measuring Devices	28

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 2 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5		Measurement, Analysis, and Improvement	N/A
5	5.1	General	29
5	5.2	Monitoring and Measurement	29
5	5.3	Internal Auditing	30
5	5.4	Monitoring and Measurement of Processes	31
5	5.5	Monitoring of Measurement of Product	31
5	5.6	Control of Non-conforming Product	32
5	5.7	Analysis of Data	34
5	5.8	Improvement	35
5	5.9	Corrective Action	36
5	5.10	Preventive Action	37

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 3 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

Section 1: Quality Management System

1.1 General Requirements

- 1.1.1** Executive Management has established and maintains a documented Quality Management System in order to meet the requirements of the ISO 13485:2003 Standard, the ISO 9001:2000 standard, Canadian CMDR requirements, the Medical Device Directive 93/42/EEC of the European Union, and the Quality System Regulations of the FDA.
- 1.1.2** As the business scope of American Diagnostic Corporation expands to encompass new countries and regions; the Quality Management System will be updated to incorporate any regulatory requirements mandated by the governments of these countries/regions. The Quality Management System will also be continually improved for effectiveness in accordance with the ISO 13485:2003 Standard, the ISO 9001:2000 standard, Canadian CMDR requirements, the Medical Device Directive 93/42/EEC of the European Union, and the Quality System Regulations of the FDA with the goal of this improvement being increased product quality, process efficiency, and customer satisfaction.
- 1.1.3** Executive Management has:
- Identified the processes needed for the Quality Management System and their application throughout the organization (as required the ISO 13485:2003 Standard, the ISO 9001:2000 standard, Canadian CMDR requirements, the Medical Device Directive 93/42/EEC of the European Union, and the Quality System Regulations of the FDA.)
 - Determined the sequence and interaction of these processes (Note: refer to the ADC process flowchart in this Quality Manual)
 - Determined criteria and methods needed to ensure that both the operation and the control of these processes are effective
 - Ensured the availability of resources and information necessary to support the operation and monitoring these processes, including the development of appropriate departments to handle defined responsibilities as part of the Quality Management System
 - Developed systems to monitor, measure, and analyze these processes
 - Implemented processes necessary to achieve planned results and developed systems to monitor and provide for continual improvement of these processes.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 4 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- 1.1.6** As new products are developed or new customer and regulatory requirements come into effect, processes will be designed to incorporate these new requirements. These new processes will fit into the process flow as illustrated above.
- 1.1.7** All processes that make up the Quality Management System are interconnected, with the output of one process becoming the input for the following process. This process flow ensures that customer and regulatory requirements are met throughout the creation of products and services.
- 1.1.8** Where processes that might affect product quality are performed through outsourcing, Executive Management has established processes to ensure that product quality is maintained and that products conform to requirements.

1.2 Documentation Requirements

- 1.2.1** The Quality Management System is documented and includes a Quality Policy and Quality Objectives, a Quality Manual, documented procedures, internal documents, and Quality Records. These documents ensure that all processes are effectively planned, implemented, and controlled.
- 1.2.2** As part of the Quality Management System documentation, all documents and files that are required by various regulatory authorities have been created and are maintained in physical or electronic form. These documents include, but are not limited to, the following:
- Design History Files (DHF)
 - Device Master Records (DMR)
 - Production History Reports
 - Product Technical Files (TFCE)
- 1.2.3** All documents and files that are required by various regulatory authorities are discussed throughout ADC's Standard Operating Procedures and work instructions. As new requirements become mandatory, additional work instructions will be created to outline the steps that ADC will take to satisfy those requirements.
- 1.2.4** Documents that make up the Quality Management System are arranged in tiers, with this Quality Manual representing the most broad, generalized document. Standard Operating Procedures are associated with this Quality Manual and make reference to it, but contain more detailed information. Work instructions provide precise, step-by-step instructions and make reference to the SOP's. This tiered structure is described in greater detail in specific work instructions.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 6 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

1.3 Quality Manual

1.3.1 This Quality Manual includes:

- The scope of the Quality Management System, including details of and justification for any exclusions,
- The documented procedures established for the Quality Management System, or references to them, and
- A description of the interaction between the processes of the Quality Management System.

1.3.2 This Quality Manual has been designed as a general guidance on, and description of, ADC's Quality Management System. It is intended for general distribution to external sources.

1.3.3 EXCLUSIONS AND NON-APPLICABLE SECTIONS

1.3.3.1 This section defines sections that ADC is excluding from the Quality Management System as well as sections that are not applicable to ADC's Quality Management System at this time. These exclusions and non-applicable sections may be included in the Quality Management System in the future should ADC's business practices or product line change in such a way that warrants their inclusion. New product designs and new processes will be evaluated by Executive Management to determine if changes to the Quality Management System would be necessary in relation to these exclusions.

1.3.3.2 The following points are excluded or non-applicable in relation to ADC's Quality Management System. A brief discussion of the justification for each exclusion is included along with the points.

- **Sterility-** As discussed in sections 4.6.2 and 4.9, ADC does not currently manufacture any devices that are provided to the customer in a sterile state. National and international requirements relating to this topic are not applicable to ADC at this time.
- **Installation and Maintenance-** As discussed in sections 4.7 and 4.8, ADC does not install any medical devices. Maintenance activities performed by ADC are limited to those activities that are performed within ADC's own facility; no maintenance activities are performed at customer owned facilities. The Quality Management System therefore will not include processes for installation or maintenance in customer owned facilities.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 7 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- **Implantable and Active Implantable Medical Devices-** As discussed in sections 4.11.5 and 5.5.4, ADC does not manufacture or distribute implantable or active implantable medical devices. Any national or international requirements for these device types are not applicable to ADC's Quality Management System at this time.

1.4 Control of Documents

- 1.4.1 Executive Management has established a document control system to ensure that documents required by the Quality Management System are controlled. They cannot be adjusted or modified by any ADC personnel without authorization.
- 1.4.2 Quality Records are maintained as indicated in specific work instructions. They are controlled and stored for at least the lifetime of ADC's products, as defined in technical files (TFCE) and other documents. This applies to obsolete revisions of documents as well as active documents that are in use.
- 1.4.3 Executive Management has established and maintains a documented procedure to define the controls needed:
- To review, approve, and ensure the adequacy of documents before their release for use
 - To update documents as they become obsolete or require correction, including a review and approval process for these updates
 - To ensure that changes and the current revision status of documents are identified
 - To ensure that documents are distributed to the appropriate personnel and that they are available as needed
 - To ensure that documents remain legible and readily identifiable
 - To ensure that documents of external origin are identified and their distribution controlled where these documents could impact the company's Quality Management System
 - To prevent the unintended use of obsolete documents by preventing their distribution, and ensuring that any obsolete documents retained for reference purposes are clearly marked "obsolete"
- 1.4.4 Document control will be performed in accordance with SOP 4.0, *Quality Management System*. This SOP makes reference to or describes the procedures used to approve and review documents, as well as discussing document retention.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 8 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

1.4.5 As part of the document control system established by Executive Management, the process for changing existing documents and the requirements for the review of these revised documents have been established to ensure that a competent review is performed that results in proper document revision.

1.5 Control of Quality Records

1.5.1 As part of the various processes and procedures that make up ADC's Quality Management System, Quality Records will be generated where appropriate to ensure that proper documentation and record keeping is performed when vital processes are completed. Quality Records are created and maintained to provide evidence of conformity to requirements and of the effective operation of the Quality Management System. These records will remain legible, readily identifiable, and retrievable.

1.5.2 Documented procedures have been established as part of the Quality Management System that define the controls needed to create and maintain Quality Records. These procedures define the processes used for identification, storage, protection, retrieval, retention time, and disposition of quality records. These procedures are included in various work instructions that apply to the departments that produce these records.

1.5.3 Quality records are controlled in accordance with SOP 4.0, *Quality Management System*. Quality Records are maintained for the lifetime of each medical device as defined in the technical documentation for that device or for a period of at least five years if the expected lifetime of a medical device is less than this time. (Note: Most medical devices manufactured by American Diagnostic Corporation have a life cycle of less than five years.)

Section 2: Management Responsibility

2.1 Management Commitment

2.1.1 Executive Management has demonstrated a commitment to the development and implementation of the Quality Management System, and continually improving its effectiveness by:

- Communicating to the Company the importance of meeting Customer as well as statutory and regulatory requirements (in particularly, requirements for safety and performance of ADC devices)
- Establishing the Quality Policy (Refer to the sections below for more details)

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 9 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- Establishing Quality Objectives that are meaningful to ADC's business model, and reviewing these objectives on a regular basis to ensure that they help to improve ADC's overall quality
- Conducting Management Reviews
- Providing the resources needed by the Quality Management System to ensure its proper operation

2.1.2 Management commitment is addressed in SOP 5.0, *Management Responsibility*, and is also documented in a variety of work instructions that form the Quality Management System.

2.2 Customer Focus

2.2.1 Executive Management will ensure that Customer requirements are determined and fulfilled with the aim of enhancing Customer Satisfaction. Through ADC's contract review process all customer orders and extended contracts are reviewed to ensure that ADC can meet all of the customer's requirements, both stated and unstated.

2.2.2 To ensure that customers are satisfied throughout the lifetime of their purchases, ADC's Customer Service Department has been established to provide customers with the services they need. Customers can purchase ADC products with the knowledge that ADC's Customer Service Department will be available with technical support, order information, and to help the customer obtain repairs as necessary.

2.2.3 Customer focus is addressed in SOP 5.0 *Management Responsibility*, and is also documented in a variety of work instructions that form the Quality Management System.

2.3 Quality Policy

2.3.1 Executive Management will ensure that the Quality Policy is appropriate to the purpose of the Company, and includes a commitment to comply with the requirements of both customers AND regulatory authorities and continually improve the effectiveness of the Quality Management System.

2.3.2 The Quality Policy will provide a framework for establishing and reviewing Quality Objectives as all Quality Objectives are designed to conform to the basic principles presented in the Quality Policy. Executive Management will ensure that the Quality Policy is communicated and understood within the Company, and will review the Quality Policy for adequacy and suitability on a regular basis.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 10 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

2.3.3 The Company Quality Policy is:

“The Company is committed to maintaining the effectiveness of, and continually improving, our Quality Management System to respond to and meet the expectations of our Customers while remaining in compliance with the legal requirements of the markets we serve.”

2.3.4 It is important to note that the above quality policy discusses *maintaining* the Quality Management System. The use of the word *maintaining* is taken to mean not only the maintenance of the existing system, but ensuring that the system is maintained in accordance with the requirements of applicable regulatory authorities. This policy ensures that ADC’s Quality Management System meets not only the requirements of our customers, but also all legal requirements in the countries where ADC conducts business.

2.3.5 In addition to the above interpretation, it is also important that the word “*customers*” is defined. When the customer is referred to in any documentation relating to ADC’s Quality Management System, any of the following could apply:

- The distributor, retailer, or reseller that directly purchased the product.
- The end user, including the doctor, nurse, and the healthcare institution using the device, or the home user.
- The patient that the device is being used on.

2.3.6 This definition of customer ensures that all safety and reliability concerns are addressed as part of ADC’s Quality Management System.

2.3.7 The Quality Policy is addressed in SOP 5.0 *Management Responsibility*, and is also documented in a variety of work instructions that form the Quality Management System.

2.4 Planning

2.4.1 QUALITY OBJECTIVES

2.4.1.1 Executive Management has established Quality Objectives at relevant functions and levels within the Company. These Quality Objectives are measurable and consistent with the Quality Policy. (Note: Quality Objectives include objectives created for ADC products that measure quality and other characteristics of ADC’s level of service.)

Note: For the latest revision of this document, refer to ADC’s Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 11 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

2.4.1.2 Quality Objectives are reviewed by Executive Management during the normal Management Review process. This ensures that progress towards these goals can be monitored, and the Quality Objectives can be revised as necessary to more appropriately match the resources and capabilities of ADC's various departments.

2.4.1.3 The Quality Objectives are addressed in SOP 5.0 *Management Responsibility*.

2.4.2 QUALITY MANAGEMENT SYSTEM PLANNING

2.4.2.1 Executive Management will ensure that the planning of the Quality Management System is carried out in accordance with the ISO 13485:2003 Standard, the ISO 9001:2000 standard, Canadian CMDR requirements, the Medical Device Directive 93/42/EEC of the European Union, and the Quality System Regulations of the FDA relevant regulatory requirements, and the Quality Policy. Executive Management will also ensure that the integrity of the Quality Management System is maintained when changes to the Quality Management System are planned and implemented.

2.4.2.2 The Quality Management System, and any changes that will be made to the Quality Management System, will be reviewed during normal Management Review Meetings prior to implementation. This will ensure that the integrity of the Quality Management System is maintained at all times.

2.4.2.3 Changes to existing processes or the creation of new processes will also be discussed during Plant Manager Meetings. These meetings provide Executive Management with an appropriate forum to discuss impending changes to the Quality Management System. Changes to existing processes can be planned out prior to implementation, with Executive Management available to provide guidance.

2.4.2.4 Quality Management System planning is performed in accordance with SOP 5.0 *Management Responsibility*.

2.5 Responsibility, Authority, and Communication

2.5.1 RESPONSIBILITY AND AUTHORITY

2.5.1.1 Executive Management has defined and documented the responsibility, authority, and the interrelation of personnel who manage, perform, and verify work within the Quality Management System. Part of this documentation includes official job descriptions, which define the various tasks performed by employees during their daily activities.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 12 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

2.5.1.2 Executive Management has developed an organization chart that depicts the quality organization. This organizational chart includes official job titles and shows the interrelation of all ADC employees. It is further addressed in SOP 5.0, *Management Responsibility*.

2.5.2 MANAGEMENT REPRESENTATIVE

2.5.2.1 Executive Management has appointed the *Quality Manager* as the Management Representative for the Quality Management System. The Management Representative has the responsibility and authority that includes:

- Ensuring that processes needed for the Quality Management System are established, implemented, and maintained by making use of resources provided by Executive Management and personnel from various departments within the company
- Reporting on the performance of the Quality Management System to Executive Management and any need for improvement (Note: this reporting is aided by the analysis of quality system data obtained through normal monitoring and measurement processes.)
- Ensuring the promotion of awareness of Customer requirements throughout the Company

2.5.2.2 Appointment of the Management Representative is addressed in SOP 5.0, *Management Responsibility*. An official job description has been created to describe all of the functions and responsibilities of the *Quality Manager* in relation to his or her duties as the Management Representative.

2.5.3 INTERNAL COMMUNICATION

2.5.3.1 Executive Management has established communication processes within the Company to ensure that processes included as part of the Quality Management System are effectively implemented and maintained. This communication ensures that the Quality Management System is effective and accomplishes stated objectives.

2.5.3.2 This communication process makes use of all computerized resources that are available at ADC, and includes the creation and continuous adaptation of an electronic Information System to provide all departments with the latest information, controlled documents, and Quality Objectives.

2.5.3.3 Responsibility, authority, and communication are addressed in SOP 5.0, *Management Responsibility*.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 13 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

2.6 Management Review

2.6.1 Executive Management will conduct a review of the Quality Management System on a periodic basis to ensure its continued suitability and effectiveness in accordance with the Quality Policy and Quality Objectives. This review will include assessing opportunities for improvement and the need for changes to the Quality Management System, including the Quality Policy and Quality Objectives. The Management Review meeting will be attended by a variety of management staff throughout ADC to ensure that all relevant information is presented and reviewed.

2.6.2 The input to Management Review will be in the form of a report. This report is prepared using quality system data collected through monitoring and measurement processes, as well as statistical data obtained during analysis. This report includes:

- Results of audits
- Customer feedback information
- Process performance and product conformity information (from a variety of sources)
- Status of preventive and corrective actions
- Follow-up actions from previous management reviews
- Planned changes that could affect the Quality Management System
- Recommendations for improvement
- New or updated regulatory requirements

2.6.3 The output from Management Review includes any decisions and actions related to:

- Improvement of the effectiveness of the Quality Management System and its processes
- Improvement of product related to Customer requirements
- Resource needs
- Corrective and preventive actions as applicable

This output will be presented in a form that is suitable for the various departments it may affect. Policies created during normal management review will be incorporated into existing work instructions and procedures where appropriate, or documented in new procedures where required.

2.6.4 Management Review is performed in accordance with SOP 5.0, *Management Responsibility*. Management Review is documented in separate work instructions that further detail the various inputs and the outputs that are achieved by this meeting.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 14 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

Section 3: Resource Management

3.1 Provision of Resources

- 3.1.1 Executive Management has determined and provided the resources needed to implement and maintain the Quality Management System and continually improve its effectiveness. These resource requirements are reviewed during normal management review to ensure that they are adequate for all of ADC's various departments.
- 3.1.2 Executive Management has provided the resources necessary to enhance Customer Satisfaction by meeting customer requirements as well as all applicable regulatory requirements for ADC's products and services.
- 3.1.3 Provision of resources is addressed in SOP 6.0, *Resource Management*, and is also documented in a variety of work instructions that form the Quality Management System.

3.2 Human Resources

- 3.2.1 Personnel performing work affecting product quality will be competent on the basis of appropriate education, training, skills, and experience. Executive Management has established a Human Resources Department to organize the process of hiring new employees.
- 3.2.2 Executive Management:
- Determines the necessary competence for personnel performing work affecting product quality (This information is included in relevant job descriptions)
 - Provides training or takes other actions to satisfy these needs
 - Evaluates the effectiveness of actions taken through regular training review
 - Ensures that its personnel are aware of the relevance and importance of their activities and how they contribute to the achievement of the Quality Objectives. This is accomplished through incentive programs, training sessions, and the appropriate communication of the Quality Policy and Quality Objectives.
 - Has established a system to maintain appropriate records of education, training, skills, and experience

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 15 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

3.2.3 Human resources are addressed in SOP 6.0, *Resource Management*. This topic is also documented in a variety of work instructions that form the Quality Management System. This documentation includes official job descriptions, which contain the educational and training requirements needed to perform specific tasks.

3.3 Infrastructure

3.3.1 Executive Management has provided and maintains the infrastructure needed to achieve conformity to product requirements. Infrastructure includes:

- Building, workspace, and associated utilities
- Process equipment, both hardware and software
- Supporting services such as transport or communication

3.3.2 As part of Executive Management's commitment to provide adequate infrastructure a dedicated IT Department has been created to safeguard all electronic data and ensure the uninterrupted operations of ADC's computer systems.

3.3.3 For those processes that affect product quality, Executive Management has established documented work instructions and procedures to ensure that any equipment that can have an impact on product quality is maintained in an adequate fashion.

3.3.4 Infrastructure is discussed in more detail in SOP 6.0, *Resource Management*.

3.4 Work Environment

3.4.1 Executive Management has established the work environment criteria needed to achieve conformity to product requirements. These criteria include procedures and processes for building cleanliness, process efficiency, and employee safety.

3.4.2 Where the health, cleanliness, or clothing of ADC personnel could adversely affect the quality and/or safety of ADC's products, criteria have been established to ensure product quality.

3.4.3 Any products produced by ADC that could be adversely affected by environmental conditions will be protected to ensure that product quality and safety does not suffer. Environmental controls necessary to safeguard these products will be documented and maintained as appropriate.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 16 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- 3.4.4** Where environmental controls are necessary, the training and skill of employees working in these controlled areas will be documented to ensure the safety of both the employees as well as ADC products.
- 3.4.5** Executive Management has established a Safety Committee to ensure that relevant employee safety issues are discussed and resolved. This committee is responsible for overseeing all aspects of employee safety, including the handling and/or storage of any hazardous materials.
- 3.4.6** As part of the procedures and documents created to control the cleanliness and work environment of ADC, contamination (with hazardous OR non-hazardous materials) of product will be controlled to prevent the spread of contamination to other products, the environment, or ADC employees.
- 3.4.7** Work environment is addressed in SOP 6.0, *Resource Management*, and is also documented in a variety of work instructions that form the Quality Management System.

Section 4: Product Realization

4.1 Planning of Product Realization

- 4.1.1** Executive Management plans and develops the processes needed for product realization. These processes include the steps necessary to design and develop new products, as well as the implementation of procedures and work instructions to detail the processes used to create and inspect the products. Planning of product realization is consistent with the requirements of the other processes of the Quality Management System and includes any services provided by ADC to our customers.
- 4.1.2** In planning product realization, Executive Management determines and documents the following, as appropriate:
- Quality Objectives and requirements for the product (In particularly regulatory requirements for the various regions that the product will be sold in.)
 - The need to establish processes, documents, and provide resources specific to the product
 - Required verification, validation, monitoring, inspection, and test activities specific to the product and the criteria for product acceptance. This includes, as appropriate, any necessary training for inspection personnel.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 17 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- Records needed to provide evidence that the realization processes and resulting product fulfill requirements

4.1.3 It is important to note that included within the requirements for a product are all technical/regulatory documentation requirements that ADC must meet in order to market a product.

4.1.4 Executive Management has established appropriate communication processes and/or committees to ensure that all relevant information relating to product development and the planning of product realization is discussed and presented to every department of ADC prior to implementation.

4.1.5 Executive Management has established appropriate risk management programs to ensure that risk assessment is included as part of the planning of product realization. Records of risk analysis are maintained as Quality Records.

4.1.6 The output from product realization planning includes all of the forms, work instructions, technical documentation, and other records associated with the development of products and processes. These materials are incorporated into ADC's Quality Management System.

4.1.7 Additional information about the planning of product realization is available in SOP 7.0, *Product Realization*.

4.2 Customer Related Processes

4.2.1 Executive Management has created a contract review system employed by ADC's Customer Service Department and sales force that ensures customer requirements are understood prior to accepting an order. The purpose of this system is to ensure order accuracy and reduce errors when customers place orders. This system also ensures that specific contracts are reviewed appropriately before acceptance. This system has been designed to determine:

- Requirements specified by the Customer, including the requirements for delivery and post-delivery activities
- Requirements not stated by the Customer but necessary for specified use or known and intended use
- Statutory and regulatory requirements related to the product
- Any additional requirements necessary to ensure the customer's satisfaction, such as providing technical information or specific documentation

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 18 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- 4.2.2** Executive Management has created a contract review system to review the requirements related to the product. This review will be conducted prior to the Company's commitment to supply a product to the Customer and will ensure that:
- Product requirements are defined and documented in physical form or using electronic systems. In most cases ADC customers will purchase existing ADC products as defined and documented in ADC's catalog, when new products are proposed, the design requirements will be documented in specific Design History Files.
 - Contract or order requirements differing from those previously expressed are resolved
 - ADC's current products (or proposed new product designs) meet the defined requirements, including the requirements for quantity as appropriate
- 4.2.3** Records of the results of contract review and actions arising from the review will be maintained as Quality Records and are available to appropriate departments so that revisions to existing contracts can be evaluated before implementation. (Ref. SOP 7.0, *Product Realization*)
- 4.2.4** In instances where a customer provides no documented statement of requirements, the customer requirements will be confirmed by ADC's Customer Service Department prior to the acceptance of any orders as part of the contract review process. Where product requirements are changed, ADC will ensure that relevant documents are amended and that relevant personnel are made aware of the changed requirements as appropriate.
- 4.2.5** As part of the process of reviewing customer requirements, ADC conducts regular reviews of catalogs and product information that is provided to the public. This review ensures a continual improvement and clarification of relevant product information.
- 4.2.6** Executive Management has implemented effective arrangements for communicating with Customers in relation to:
- Product information, including technical details about ADC's products, sales information, and service information.
 - Inquiries, contracts, or order handling, including amendments
 - Customer feedback, including customer complaints

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 19 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

4.2.7 Executive Management has established a documented system to ensure that customers are aware of advisories or recalls that apply to ADC products. This ensures end-user safety by allowing for the quick distribution of information relating to problems associated with a particular device or batch of devices.

4.2.8 Customer related processes are further addressed in SOP 7.0, *Product Realization*.

4.3 Design and Development

4.3.1 Executive Management plans and controls the design and development of products. During the design and development planning, various departments within the company will determine the following:

- The design and development stages (including the departments and resources that will be required for each stage)
- The review, verification, and validation that are appropriate to each design and development stage
- The responsibilities and authorities for design and development

4.3.2 Executive Management has established a system that manages the interfaces between different groups involved in design and development to ensure effective communication and clear assignment of responsibility for various design related activities. Planning output will be updated, as appropriate, as the design and development progresses.

4.3.3 All stages of the design and development process are recorded in relevant Design History files, ensuring that a complete history of a product's development is available for review.

4.3.4 During the planning phase of product development, inputs related to product requirements will be determined. These inputs will form the basis for a Design History File (DHF) for that design. These Design History Files will become the main record of all product development performed at later stages of the design process. (Ref. SOP 7.0, *Product Realization*.) The DHF will include:

- Functional, performance, and safety requirements as applicable
- Any regulatory requirements that apply to the product
- Information relating to existing designs that are similar to the new product
- Other requirements essential for design and development, including ADC specifications or components lists as applicable

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 20 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- Risk assessment information as applicable (Note: risk assessment information may be contained within product specific technical files as well as/instead of product Design History Files.)

- 4.3.5** Inputs of the design development process will be reviewed after they have been determined. Throughout the process of product development this information is reviewed by relevant departments to ensure that the implementation of new processes during product realization meets all stated requirements.
- 4.3.6** Where requirements are found to be incomplete, ambiguous, or in conflict with one another, these requirements will be clarified and corrected before any product is manufactured or processes are implemented.
- 4.3.7** Design and development output will include all relevant information necessary to verify that the new product meets all stated requirements. Through the design and development system established by Executive Management, this verification is performed and the results are approved and recorded.
- 4.3.8** In addition to design verification, the design and development system also ensures that product designs are validated during the final stages of the development process. This validation ensures that the final product is able to meet all of the performance and safety requirements for the intended use/application of the product. Records of this validation are maintained as part of a Design History File.
- 4.3.9** Validation of a product design may include clinical evaluations as required by national or international standards.
- 4.3.10** The outputs from the design and development system include relevant work instructions, forms, procedures, and other documentation to ensure that purchasing, production, customer service, quality control (product acceptance criteria), and service processes are established. These outputs will also include documentation related to the safety and performance characteristics of the product where appropriate. All design output records are included in product Design History Files.
- 4.3.11** During the course of design development, reviews of existing information about a new design will be discussed and reviewed with appropriate departments to ensure that any problems are identified and corrected so that the product can meet all stated requirements. Executive Management has established a design communication system that includes checklists and committee meetings to ensure designs are thoroughly reviewed.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 21 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- 4.3.12** The review of a new design will include representatives from Executive Management as well as members of ADC's design team to ensure that all aspects of the design are properly communicated to those departments that require specific information. Records of all design reviews are included in Design History Files.
- 4.3.13** As part of ADC's design and development process, changes to existing designs are reviewed, verified, and validated before they are approved and implemented. This review will ensure that the finished product, as well as components for the product will continue to function as intended. All reviews are conducted in the same fashion as reviews for new product designs.
- 4.3.14** All stages of the design development process are documented within individual Design History Files. Additional procedures and processes for product design and development can be found in SOP 7.0, *Product Realization*, and throughout ADC's Quality Management System.

4.4 Purchasing

- 4.4.1** Executive Management has established a Quality Control Department to ensure that purchased product conforms to specified purchase requirements. All products and components arriving at ADC are inspected and tested to verify the quality of each shipment.
- 4.4.2** Executive Management has established a documented sampling plan to ensure that inspected items are randomly sampled during the inspection process. This sampling plan is based upon supplier performance and item or component impact on product realization or the final product.
- 4.4.3** Executive Management has established a supplier survey and evaluation system to ensure that supplier quality is monitored and measured. This system includes documented procedures for supplier selection, evaluation, and re-evaluation. Records of the results of evaluations and any necessary actions arising from the evaluation are maintained as part of this system.
- 4.4.4** Additional information relating to supplier selection and evaluation can be found in SOP 7.0, *Product Realization*.
- 4.4.5** Executive Management has established the Purchasing Department to purchase items and components used in ADC products. The Purchasing Department ensures that all relevant purchasing information is described in the documentation supplied to the vendor. This information includes:

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 22 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- Product identification, including quantity and price
- Requirements for product approval
- Requirements for the approval of a vendor's procedures, processes, and/or equipment
- Training requirements applicable to product related processes
- Quality Management System requirements if applicable
- Traceability requirements necessary for the purchased product or components (where applicable)

4.4.6 Purchase requirements are verified by the Purchasing Department prior to their communication to ADC's vendors. In instances where ADC will perform verification of purchased products at a vendor's premises, the details of this verification will be included in the purchasing information supplied to the vendor. Records of this type of verification are maintained as part of ADC's Quality Records.

4.4.7 Additional procedures and processes related to ADC's Purchasing Department and product verification activities can be found in SOP 7.0, *Product Realization*.

4.5 Product and Service Provision

4.5.1 Executive Management has established the processes and procedures necessary to plan and carry out production and service processes under controlled conditions. These processes and procedures are interrelated, as described in the basic ADC process flowchart presented in this Quality Manual. Controlled conditions may include, but are not limited to, the following (as applicable):

- The availability of information that describes the characteristics of the product, including relevant documentation required by national or international regulations (such as Device Master Records or Technical Files)
- The availability of work instructions to the employees that require them, including reference materials and/or measurement standards
- The requirements for equipment where that equipment could affect the performance or safety of the product or the safety of ADC personnel
- The availability and use of monitoring and measuring devices, particularly for the testing and/or verification of medical devices that have a measurement function
- The implementation of monitoring and measurement, including the inspection of incoming materials, in process products, and finished products

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 23 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- The implementation of release, delivery, and post-delivery activities
- Labeling and packaging requirements and any special environmental controls or employee training necessary to ensure product conformity

4.5.2 Executive Management has established the necessary processes to record traceability information related to ADC products on the market to ensure that safety or performance issues can be appropriately addressed. These records include production history and inspection status information and are verified at the time of their creation.

4.5.3 Additional information relating to the control of production and service provision can be found in SOP 7.0, *Product Realization*, and is documented in various work instructions and procedures that make up ADC's Quality Management System.

4.6 Cleanliness of Product and Contamination Control

4.6.1 Executive Management has established environmental controls to ensure that the cleanliness of ADC products is maintained throughout the manufacturing and storage processes. These controls ensure that products with special cleanliness requirements are handled appropriately to avoid contamination of new or existing products.

4.6.2 As part of the design and development process, the cleanliness requirements for new products will be evaluated prior to any production of these products. Where new environmental controls are needed, Executive Management will ensure that these controls are in place before any products are produced. These controls may include environmental controls for sterility of products if applicable to a new product's design, although ADC currently does not produce sterile products.

4.6.3 Additional information relating to product cleanliness and contamination control may be found in SOP 6.0, *Resource Management*, and in various work instructions and procedures that make up ADC's Quality Management System.

4.7 Installation Activities

4.7.1 ADC does not currently produce products that require special installation processes. As part of the design and development processes, new products that are developed will be evaluated in order to determine whether or not installation will be necessary for the end-user.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 24 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

4.7.2 Should it be determined that a new product produced by ADC will require additional installation processes, these processes will be clearly documented, including acceptance criteria as well as verification activities to ensure that the installation of the medical device was completed correctly.

4.7.3 Records of any installation processes will be maintained as Quality Records by ADC.

4.8 Servicing Activities

4.8.1 Executive Management has established the Service Department to ensure that ADC products are properly serviced and maintained. These servicing activities occur on ADC's premises only, and at this time do not include any repairs or maintenance at customer owned facilities.

4.8.2 The Service Department has a number of established procedures and work instructions to ensure that ADC products are repaired to acceptable levels of performance and safety. Additional information relating to these procedures may be found in SOP 7.0, *Product Realization*.

4.8.3 Records of repair/maintenance work performed by the Service Department are maintained as Quality Records.

4.9 Particular Requirements for Sterile Medical Devices

4.9.1 At this time, ADC does not produce sterile medical devices. Should a product be developed that requires sterilization by ADC, the requirements for sterility will be clearly documented, with process parameters clearly defined and sterilization records maintained that are traceable to each batch of product produced. Verification and validation of these sterilization processes would be included in the documented procedures and work instructions.

4.10 Validation of Processes for Product and Service Provision

4.10.1 Although most processes performed by ADC to manufacture finished products can be verified by subsequent monitoring or measurement, any process that is established for production purposes that cannot be verified will be validated prior to its acceptance as a standard manufacturing process.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 25 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

4.10.2 This validation includes any processes where deficiencies become apparent only after the product is in use or a service has been delivered. This validation will demonstrate the ability of the processes to achieve planned results. Records of any validation will be maintained as part of ADC's Quality Management System.

4.10.3 Where production and service processes do not have easily verifiable outputs, these processes are reviewed and approved periodically. This review includes a review of equipment, personnel, record keeping, and specified procedures used in various departments at ADC. This review will allow these processes to be revalidated where applicable.

4.10.4 If a process is created utilizing computer software that could affect the ability of the product to conform to requirements, this software will also be validated prior to its use for manufacturing processes.

4.11 Identification and Traceability

4.11.1 Executive Management has established an item numbering system to ensure that products, sub-assemblies, and component parts are easily identified throughout the product realization process.

4.11.2 Executive Management has established inspection status procedures to ensure that products are identified with respect to monitoring and measurement requirements. This includes all aspects of the inspection process, from received goods to finished products. These procedures ensure that only products that have passed through specific monitoring and measurement processes are provided to the customer.

4.11.3 Executive Management has established a lot numbering system to ensure product traceability. While traceability is not a requirement for any of ADC's existing products, it allows ADC to easily recall potentially non-conforming products from the market. This lot numbering information is recorded and maintained as part of ADC's Quality Records.

4.11.4 Executive Management has established service and repair processes to ensure that returned items are clearly identifiable and are not confused with conforming product. Additional processes also ensure that non-conforming components and other materials are not used in manufacturing processes.

4.11.5 ADC currently does not produce active implantable medical devices or implantable medical devices with additional record keeping requirements.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 26 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

4.11.6 Additional information about product identification and traceability is available in SOP 7.0, *Product Realization*, and is also documented in various procedures and work instructions that make up ADC's Quality Management System.

4.12 Customer Property

4.12.1 Care will be exercised with Customer property while it is under ADC's control or being used by ADC. Customer property for use or incorporation into finished products will be identified, verified, protected, and safeguarded.

4.12.2 Any Customer property that is lost, damaged, or otherwise found to be unsuitable for use will be reported to the Customer and records shall be created to document such issues. Refer to SOP 7.0, *Product Realization* for additional information about procedures dealing with customer property.

4.12.3 It is important to note that "customer property" can be interpreted to mean intellectual property or confidential health information as well as physical property.

4.13 Preservation of Product

4.13.1 All products and components will be preserved throughout the manufacturing process to ensure product conformity with specifications. These preservation processes will be in effect up to and including the time of delivery to the intended destination.

4.13.2 Preservation of product includes all procedures and work instructions necessary to describe the identification, handling packaging, storage, and protection requirements for ADC products. Where applicable, these processes also include environmental controls to ensure that factors in the environment do not degrade the quality of ADC products or components.

4.13.3 Executive Management has established processes to ensure that products with limited shelf lives or components with limited shelf lives are monitored to ensure that expired products or components are not used in manufacturing processes.

4.13.4 Additional information relating to preservation processes can be found in SOP 7.0, *Product Realization*, and is also available in various work instructions and procedures that make up ADC's Quality Management System.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 27 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

4.14 Control of Monitoring and Measuring Devices

- 4.14.1** Executive Management has established the necessary monitoring and measurement processes needed to ensure that products and components conform to specified requirements. These processes include detailed instructions for the use of test equipment to verify product conformity and, where necessary, verification of the environmental conditions necessary to perform testing. These processes ensure that monitoring and measurement activities can and are carried out to meet all stated requirements.
- 4.14.2** As part of the processes established to control measuring equipment, this equipment is calibrated or verified at specific intervals against measurement standards traceable to the National Institute of Standards and Technology (NIST). Procedures have been created to ensure that all equipment is within calibration prior to use. Where no standards exist to indicate the calibration requirements for measuring equipment, the basis used for calibration or verification is recorded in various work instructions or procedures.
- 4.14.3** Measuring equipment will be adjusted or re-adjusted as necessary to ensure that it is within acceptable calibration prior to use. In addition, all equipment will be identified to enable the calibration status to be determined, safeguarded from adjustments that would invalidate the measurement result, and protected from damage and deterioration during handling, maintenance, and storage.
- 4.14.4** Executive Management has established procedures and processes to ensure that measurement results that were based on non-conforming (out of calibration) equipment are assessed and the validity of these results are recorded to ensure that product quality and customer safety has not deteriorated as a result of the non-conforming equipment. This system includes the process of correcting non-conforming product and ensuring customer satisfaction.
- 4.14.5** As part of the processes used to control monitoring and measurement equipment, Executive Management has established a Calibration Database to ensure that all calibrated and controlled equipment is recorded, and the results of calibration and/or verification are maintained as Quality Records.
- 4.14.6** Included within ADC's process for controlling monitoring and measurement equipment is any computer software used for these purposes. This software would require verification in the same fashion as physical equipment.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 28 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- 4.14.7 Additional information about ADC's monitoring and measurement equipment can be found in SOP 7.0, *Product Realization*, and in various work instructions and procedures that make up ADC's Quality Management System.

Section 5: Measurement, Analysis, and Improvement

5.1 General

- 5.1.1 Executive Management plans and implements the monitoring, measurement, analysis, and improvement of processes needed to demonstrate product conformity, ensure conformity of the Quality Management System, and to maintain the effectiveness of the Quality Management System.
- 5.1.2 A variety of procedures and processes have been established to document this improvement and maintenance process. The creation of these procedures and processes includes the determination of applicable methods, including statistical techniques, necessary to maintain and improve the Quality Management System, and the extent of their use.
- 5.1.3 Refer to SOP 8.0, *Measurement, Analysis, and Improvement*, for additional information relating to these processes.

5.2 Monitoring and Measurement

- 5.2.1 As part of ADC's Quality Management System, several systems have been established by Executive Management to ensure that product and customer related information from the field is recorded, reviewed, and acted upon.
- 5.2.2 An electronic Complaint System has been established to monitor customer feedback and to record customer complaints. This system has been designed to ensure that all customer complaints are investigated in a timely fashion and that the results of this investigation are recorded as part of ADC's Quality Records.
- 5.2.3 In addition to this complaint system, processes have been established that allow ADC to monitor customer feedback and customer satisfaction levels. This feedback is analyzed by Executive Management to allow ADC to improve its Quality Management System.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 29 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5.2.4 Executive Management has direct overview of both the Complaint System and the feedback processes. This allows any quality issues or potential safety problems to be addressed in a timely fashion through ADC's corrective/preventive action system. Complaints and feedback form an early warning system about product issues.

5.2.5 ADC's Quality Management System will make use of experience gained from post-production feedback to ensure that our products continue to meet all requirements, and to ensure that our level of service continues to satisfy our customers.

5.2.6 Additional information relating to ADC's monitoring and measurement system can be found in SOP 8.0, *Measurement, Analysis, and Improvement*, and in various work instructions and procedures that make up ADC's Quality Management System.

5.3 Internal Auditing

5.3.1 Executive Management has established an internal auditing process to audit the Quality Management System. This auditing process ensures that the Quality Management System has been effectively implemented and is maintained. Through continuous monitoring of internal processes, the auditing process will also:

- Ensure that ADC is compliant with the ISO 13485:2003 Standard, the ISO 9001:2000 standard, Canadian CMDR requirements, the Medical Device Directive 93/42/EEC of the European Union, and the Quality System Regulations of the FDA.
- Ensure that ADC is compliant with national quality management system requirements
- Ensure that ADC meets the requirements of the Medical Device Directive as it applies to the Quality Management System
- Ensure that skills, training, and/or departmental performance does not degrade by re-auditing departments on a pre-scheduled basis

5.3.2 As part of ADC's internal auditing process, audits will be conducted for every department within the company. As each audit is planned the importance of the processes being reviewed will be evaluated to ensure that audits are conducted appropriately. Auditors will be selected to ensure that all audits are conducted in an impartial fashion, and so that auditors will not audit their own work.

5.3.3 ADC's auditing process includes work instructions and procedures that define audit frequency, scope, and auditing methods. This documentation also defines the responsibilities of planning and conducting audits, as well as the reporting requirements.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 30 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5.3.4 Departments that have non-conformities detected during the internal auditing process will correct those non-conformities in accordance with ADC's established corrective/preventive action system. These non-conformities will be corrected in a timely manner (including the underlying cause of the non-conformity). Follow-up audits and/or inspections as required by corrective/preventive actions will verify that corrections have been made and that they are effective in eliminating the detected non-conformities.

5.4 Monitoring and Measurement of Processes

5.4.1 Executive Management has established various monitoring and measurement techniques to ensure that the processes associated with the Quality Management System are monitored. This ensures that processes can achieve planned results.

5.4.2 These monitoring and measurement techniques are reviewed regularly with management staff so that all departments are aware of their status in terms of quality and the achievement of planned objectives. Issues that are discovered during this review of process data allows corrective and preventive actions to be taken to ensure that there is no adverse impact on product conformity.

5.5 Monitoring and Measurement of Product

5.5.1 Products manufactured by ADC are monitored and measured to ensure that all products conform to documented specifications. As determined by Executive Management, this monitoring and measurement is performed at three stages of product realization; incoming inspections, in-process inspections, and finished product inspections.

5.5.2 All monitoring and measurement processes result in Quality Records that provide evidence of product conformity to acceptance criteria. These records indicate the date of the inspections and the person(s) authorizing the release of finished product as applicable. The requirements for these records are included in individual work instructions that detail specific inspections performed throughout ADC's Quality Management System.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 31 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- 5.5.3** Products are not released and shipped to customers until these inspections have been completed and all products have been determined to comply with the stated requirements. Where products have been produced that do not meet all of the requirements that were originally stated, these products will only be authorized for release by Executive Management and, where appropriate, by the customer. (Note: the concept of releasing “products that do not meet stated requirements” is taken to mean products that have cosmetic defects only; it doesn’t refer to non-conforming product that could affect the health or safety of the end-user or patient.)
- 5.5.4** At this time, ADC does not currently manufacture implantable or active implantable medical devices that would require additional record keeping during the inspection/testing activities prior to product release. Even without the production of these products, as part of the record keeping processes that make up ADC’s Quality Management System, the identity of the personnel performing inspection and testing activities is recorded during IQC and FQC inspections.
- 5.5.5** Additional information relating to ADC’s inspection and testing processes can be found in SOP 8.0, *Measurement, Analysis, and Improvement*, and in various work instructions and procedures that make up ADC’s Quality Management System.

5.6 Control of Nonconforming Product

- 5.6.1** Executive Management has established documented procedures and work instructions to control the disposition, retention, and marking of product that does not conform to requirements. This system is designed to ensure that non-conforming products or components are not used in finished products or delivered to customers.
- 5.6.2** The departments responsible for handling non-conforming products or components, as well as the controls used to identify these materials and to quarantine them, are included in various work instructions and procedures that make up ADC’s Quality Management System.
- 5.6.3** Nonconforming product is dealt with in one or more of the following ways:
- By taking action to eliminate the detected nonconformity, either through the re-work of affected materials or the replacement of affected components/products
 - By authorizing its use, release, or acceptance under concession by Executive Management and, where applicable, by the Customer

Note: For the latest revision of this document, refer to ADC’s Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 32 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

- By taking action to preclude its original intended use or application through reject marking procedures and quarantine of affected stock

5.6.4 Due to the fact that ADC products are medical devices, non-conforming product will be accepted for use through concession only after it has been determined that all applicable regulatory requirements for the product have been met. (In other words, while a non-conforming product with a cosmetic defect might be accepted for use through concession, a non-conforming product that did not meet applicable safety requirements would not.)

5.6.5 Records of the nature of nonconformities and any subsequent actions taken, including concessions obtained, are maintained as part of ADC's Quality Management System. These records include the identity of any personnel authorizing the use of non-conforming product (with the customer's concession where appropriate).

5.6.6 As part of ADC's system for handling non-conforming materials, any materials that are re-worked to make them conforming to ADC requirements will be subjected to re-inspection to ensure that the re-work has succeeded in correcting the detected non-conformities.

5.6.7 ADC's Complaint System, as well as customer feedback and information obtained from the Service Department are used to help detect any potential non-conforming product in the field. When nonconforming product is detected after delivery or use has started, appropriate actions to the effects, or potential effects, of the nonconformity will be taken through ADC's corrective/preventive action system. These actions will be documented and records will be included as Quality Records.

5.6.8 As part of the process of handling non-conforming materials, materials that require re-work will be re-worked in accordance with established work instructions and procedures. If a documented work instruction is not adequate to cover the processes required to adequately re-work a product, a new work instruction will be created (with the same authorization and approval process as other work instructions) to document the re-work process. This work instruction that details the rework will be evaluated to ensure that the rework process does not adversely affect the finished product. (Note: The rework process may also be defined in associated NCMR or CAPA files as appropriate.)

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 33 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5.6.9 Additional information relating to the control of nonconforming product may be found in SOP 8.0, *Measurement, Analysis, and Improvement*, and in various work instructions and procedures that make up ADC's Quality Management System.

5.7 **Analysis of Data**

5.7.1 As part of the Quality Management System, Executive Management has established processes for the collection and analysis of quality related data. (Note: this data also includes relevant customer satisfaction information.)

5.7.2 The data collected is used to determine the suitability and effectiveness of ADC's Quality Management System and to note areas where improvement can be made. This data is reviewed by Executive Management and specified department supervisors to ensure that trends and quality issues detected during analysis is properly communicated and addressed. This data is presented in several different forums, including, but not limited to:

- Regularly scheduled Management Review Meetings
- Plant Manager Meetings (including plant departmental supervisors)
- New Product Meetings (focusing on new product development or existing product issues)

5.7.3 The data analyzed as part of ADC's Quality Management System includes, but is not limited to:

- Customer satisfaction (including customer feedback)
- Conformity to product requirements
- Characteristics and trends of processes and products, including opportunities for preventive action (Note: process analysis includes all aspects of ADC's product realization cycle as well as after-market services)
- Supplier evaluations

5.7.4 Analysis of data is performed in accordance with SOP 8.0, *Measurement, Analysis, and Improvement*. Additional information relating to this process can also be found in various procedures and work instructions that make up ADC's Quality Management System.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 34 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5.8 Improvement

5.8.1 The purpose of ADC's Quality Management System, beyond the basic concept of product quality and customer satisfaction, is to ensure that ADC continues to grow as an organization and improve its operations both to increase efficiency, and to increase product quality and customer satisfaction levels.

5.8.2 ADC's Quality Management System incorporates various different tools that help make this continual improvement effort possible. Some of these tools are detailed below:

- ADC's Quality Policy has been designed to outline the basic quality goals that the company hopes to achieve
- ADC's Quality Objectives outline more specific improvement objectives, and are reviewed by Executive Management to ensure that they are suitable for the company and to gauge their effectiveness
- ADC's internal quality auditing system ensures that departments are performing processes as determined by written work instructions and other documents, areas of concern are noted and corrected
- Data analysis allows for trending of potential quality issues and customer satisfaction levels
- ADC's corrective/preventive action systems allow issues to be averted before a quality issue develops, or to correct existing problems and improve efficiency
- The Management Review process allows Executive Management to review the whole Quality Management System and make improvements where appropriate.

5.8.3 As part of ADC's commitment to quality, a system has been implemented to allow ADC to issue advisory notices or recalls on ADC products that have been supplied to our customers. This system can be implemented whenever necessary to ensure that customers are aware of any urgent quality issues that relate to ADC products. This system is further supplemented by ADC's Customer Service and Quality Assurance Departments, which provide customers with up-to-date, accurate technical and/or regulatory information as necessary.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 35 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5.8.4 ADC's continual improvement initiative includes a Complaint System. This system is designed to record the details of all customer complaints, and allows for this information to be routed to the departments responsible for the complaint so that investigation may be performed. All complaints are investigated and reviewed so that the customer's satisfaction is guaranteed, and the actions taken to resolve the complaint (and underlying issues relating to the complaint where appropriate) will be documented so that they can be easily reviewed.

5.8.5 Where investigation indicates that a customer complaint may have been caused by activities that are outside of ADC's direct control, ADC's advisory notice system ensures that this information is shared with all relevant parties, allowing for the resolution of any quality issues.

5.8.6 ADC's Quality Management System includes a process for notifying relevant regulatory authorities when adverse events occur that involve ADC devices. The criteria for reporting these events and the method of reporting are included in documented work instructions and procedures.

5.9 Corrective Action

5.9.1 Executive Management has established corrective action systems to ensure that corrective actions are issued and implemented to effectively handle customer complaints and product non-conformities, investigate the causes of non-conformity, record the results if the investigation determines action(s) must be taken to alleviate the problems, and devise methods for ensuring corrective action(s) are effective. (It is important to note that ADC's Complaint System is a stand-alone corrective action system specifically for dealing with customer complaints.)

5.9.2 ADC's corrective action system includes processes for reviewing nonconformities and determining their cause. Once this determination has been made, actions to correct the nonconformity can be developed. Actions are also developed to prevent a recurrence of the nonconformity in the future. These actions are then implemented, and are available to Executive Management for review at any time.

5.9.3 As part of the implementation of corrective actions, it may be necessary to update existing Quality Management System documentation to prevent future recurrences of the detected non-conformity. Updates to existing documentation that are required during corrective actions are performed in the same fashion as regular document revisions.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 36 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		



Quality Manual

5.9.4 Corrective action files will contain records detailing the entire corrective action process, from the identification of the problem and root cause analysis to the details of the actions taken to correct the problem including the results of any investigations performed. These records allow for the review of the corrective action files before they are closed to ensure that they have been effective.

5.9.5 Additional information relating to the corrective action system can be found in SOP 8.0, *Measurement, Analysis, and Improvement*, and in various work instructions and procedures that make up ADC's Quality Management System.

5.10 Preventive Action

5.10.1 Executive Management has established a preventive action system to ensure that preventive actions will be issued and implemented to detect, analyze, and eliminate potential causes of problems. Actions will include the steps to implement preventive actions, methods to ensure implemented actions are effective, and procedures to maintain records of all preventive actions for Management Review.

5.10.2 As with the corrective action system, the preventive action system will involve the creation of preventive action files that contain all relevant information about specific preventive actions. These files will include the initial potential problem, root cause analysis, and investigation results, and will allow a review of the preventive action file before it is closed to ensure that the preventive action has been successful.

5.10.3 The preventive action system includes processes for:

- Determining potential nonconformities and their causes
- Evaluating the need for action to prevent occurrence of nonconformities,
- Determining and implementing action needed
- Recording the results of action taken
- Reviewing preventive action taken

5.10.4 Additional information relating to the preventive action system can be found in SOP 8.0, *Measurement, Analysis, and Improvement*, and in various work instructions and procedures that make up ADC's Quality Management System.

Note: For the latest revision of this document, refer to ADC's Information System

Filename:	Quality Manual	Created on :	1/6/03	Revision:	5	Created By:	M. Falco
Page:	Page 37 of 37	Revision Date:	8/22/08	Approved By:	K. Bourne		